

Quality Control Test Certificate

Product name	SureSeq MyPanel Custom Myeloid – 49 Gene Plus Panel
Product no.	602017-16/96
Batch no.	5015519-2
Certificate issue no.	1

Acceptance Criteria

Number of samples per MiSeq lane: 16

DNA input: 1000 ng

SureSeq MyPanel Custom Myeloid – 49 Gene Plus Panel	Specification	Pass/Fail
Mean Target Coverage:	>1000x	PASS
Minimum Target Coverage: 99.0% Target Bases Covered at:	>500x	PASS
% Target Bases Covered at >20% Mean:	99.0%	PASS

This product has been approved for release by:

Name: Vincent Wong Position: Quality Associate Date: 06 February 2019

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.