

Quality Control Test Certificate

| | |
|-----------------------|--------------------------------|
| Product name | CytoSure Genomic Labelling Kit |
| Part no. | 020020 |
| Batch no. | 16831284 |
| Certificate issue no. | 1 |

Acceptance Criteria

| Test | | Specification | Pass/fail |
|---|-----|---------------|-----------|
| Dye incorporation | Cy3 | >250 pmol | PASS |
| | Cy5 | >250 pmol | PASS |
| DNA yield | | >6µg | PASS |
| Array performance (Derivative Log Ratio Spread) | | <0.25 | PASS |
| Date 22 nd June 2018 | | | |
| Initials DH | | | |

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2008 certified management system.

Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.